

Table D10-2 Gamma Exposure Rate Differences of Two NaI Detector Heights

Three-Foot NaI Exposure Rate ($\mu\text{R/hr}$)	Corresponding Predicted 4.5-Foot NaI Exposure Rate ($\mu\text{R/hr}$)	Difference Between the Three-Foot and 4.5-Foot NaI Exposure Rates	
		($\mu\text{R/hr}$)	(Percent)
25	24.9	0.10	0.4
30	29.0	1.0	3.3
35	33.1	1.9	5.4
40	37.2	2.8	7.0
45	41.3	3.7	8.2
50	45.4	4.6	9.2